


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 10 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130191
Name		J Swartz	ID# 4314 Date 5/8/09
A	Agency	SCDL	Phone # 269-5592
Instrument Location		SCDL	
Alco S/N		SC0108	Target Value .080 High Pressure 800 PSI
B	Alco Test Values	.061 1 st Alco	.082 2 nd Alco
Signature			
(OVER)			

[Handwritten initials]
9/10/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Oct, 2009.


Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DataMaster cdm 130191

MAY 08, 2009

OPERATOR'S NAME:

SWARTZ/J

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

L

O.L. #: L

DEPT/AGENCY: AAA1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: SCOTT08

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.27 in
ALCO TARGET .000 16:26
BLANK TEST .000 16:27
INTERNAL STANDARD VERIFIED 16:27
ALCO TV 30.27 in .001 16:27
BLANK TEST .000 16:28
SUBJECT SAMPLE .000 16:29
BLANK TEST .000 16:29
ALCO TV 30.27 in .002 16:30
BLANK TEST .000 16:30

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DataMaster cdm 130191

MAY 08, 2009

TIME 16:24

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.27 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~